

Attach Recorder Chart Here

DAILY TEMPERATURE DOCUMENTATION FOR RED BLOOD CELL STORAGE

Hospital ID:

Blood Storage Unit:

Recorder Chart Changed: Day: _____ Date: _____ Time: _____ By: _____

Day	Date	Time	Recorder Temperature °C	Manual/Digital Thermometer °C		A/U	Tech	Check Inventory Level
				Upper	Lower			

NOTE: Acceptable storage temperature for Red Blood Cells is 1 to 6 °C.

Recorder chart and upper and lower thermometer reading allowances should be determined by each facility but should be within a maximum of 2°C from each other.

KEY: A = Acceptable U = Unacceptable

Comments: _____

Transfusion Service Supervisory Review By: _____ Date: _____